

Preventative Maintenance Checklist

Frequency	Preventative Maintenance	ltem	Completed by	Date	Order Link
HOT before cleaning it CAUTION! Weekly Daily Bi-Weekly	Replace distilled water in reservoir as needed Clean door gasket with mild detergent, water, and soft cloth/sponge. Check visually that gasket is intact, not loose, and clean. Wipe surrounding area with Optim disinfecting wipes.				Optim1 6"x7" Wipes
	If autoclave is only used occasionally, drain water from mineral free water reservoir once a week and refill with fresh mineral-free water or distilled water. Weekly or if text alarm of "full waste water reservoir" appears (whichever comes first) drain water from the waste water reservoir. Clean the outer parts of the autoclave with a soft cloth. Once Every 2 Weeks Perform "System Clean Process."	STER CLASS THE CLASS THE C			Image: Second
autoclave IS NOT Monthly	Clean and descale the chamber. Clean the outer parts of the autoclave with a soft cloth. Clean the drain filter of the autoclave.				
- <mark>Make sure the a</mark> u Periodic Maintenance	 Every 3 Months: Check door seal for physical damage and replace if there is damage or leakage. Every 6 Months or 1000 cycles: (Whichever comes first:) Replace the air filter. Every 6 Months or 1000 cycles: (Whichever comes first:) Replace the door seal. 				I - Top bor Gasket RJJ. N22-4029
CAUTION! - Yearly	Annual PM - Replace Door Gasket, Air Filter, and perform "System Clean Process" with TSC Cleaning Tablets *If you have questions please call our FREE Technical Support Line				Tuttnauer T-Top 10 PM Kit



Tuttnauer T-Top 10



StatimUSA.com - Free Technical Support Line 704-966-1650 Option 3

FREE TECH SUPPORT