



Sterilizer Autoclave Solutions

INFECTION CONTROL SPECIALISTS

Preventative Maintenance Checklist

Frequency	Preventative Maintenance	Item	Completed by	Date	Order Link
Daily	Replace distilled water in reservoir as needed				 Optim1 6"x7" Wipes
	Clean door gasket with mild detergent, water, and soft cloth/sponge. Check visually that gasket is intact, not loose, and clean. Wipe surrounding area with Optim disinfecting wipes.				
Weekly Bi-Weekly	If autoclave is only used occasionally, drain water from mineral free water reservoir once a week and refill with fresh mineral-free water or distilled water.				  Tuttnauer i-System Clean Tablets How to Run "System Clean Process"
	Weekly or if text alarm of "full waste water reservoir" appears (whichever comes first) drain water from the waste water reservoir.				
	Clean the outer parts of the autoclave with a soft cloth.				
	Once Every 2 Weeks Perform "System Clean Process."				
Monthly	Clean and descale the chamber.				  T-Top 11 Chamber Descaling (B01P22475) T-Top 11 Chamber Descaling Process (B01P22475)
	Clean the outer parts of the autoclave with a soft cloth.				
	Clean the drain filter of the autoclave.				
Periodic Maintenance	Every 3 Months: Check door seal for physical damage and replace if there is damage or leakage.				  T-Top 11 Door Gasket OE (RU1424-024) T-Top 11 Air Filter OE (SUS-575-010)
	Every 6 Months or 1000 cycles: (Whichever comes first:) Replace the air filter.				
	Every 6 Months or 1000 cycles: (Whichever comes first:) Replace the door seal.				
Yearly	Annual PM - Replace Door Gasket, Air Filter, and perform "System Clean Process" with TSC Cleaning Tablets				 T-Top 11 OEM PM KIT
	*If you have questions please call our FREE Technical Support Line				



Tuttnauer T-Top 11

StatimUSA.com - Free Technical Support Line 704-966-1650 Option 3



FREE TECH SUPPORT